

SPEC-DEMO Workshop

Internal calibration record

SPEC-DEMO-001-1

Internal record of a calibration. This is not an accredited calibration certificate.

Instrument and calibration

Record	SPEC-DEMO-001-1
Calibrated on	2026-10-30
Control number	SPEC-DEMO-001
Type	Digital caliper
Serial number	DEMO-101
Manufacturer / model	SPEC-DEMO · 150 mm
Location	Inspection bench
Performed by	SPEC-DEMO Metrology (external provider)
Certificate reference	SPEC-DEMO-REF-001

Results

As found	Fail
Adjusted	Yes
As left	Pass
Repaired	No
Fit for purpose	Yes
Final result	Pass

Measurements

Parameter	Nominal	Limits	As found	As left	Reference standard	Result
Length	50 mm	49.98 – 50.02 mm	50.04 mm · Fail	50.01 mm · Pass	—	Pass

Out of tolerance

Status	Closed
Product impact assessment	Fictional batch rechecked; no affected product identified.
Corrective action	SPEC-DEMO adjustment and recheck
Closing decision	Fictional reinspection complete. Instrument returned to service.
Closed by	SPEC-DEMO Alex on 2026-10-31 12:00

Statement

Fictional demonstration data - not for operational use.

Performed by

Name

Signature

Date

Reviewed by

Name

Signature

Date